

Annexure - I

Dear Sir,

This is with reference to your mail dated 10/12/25 regarding clarification on certain points. Our committee has gone through the points raised by you and we have tried to address it as mentioned below:

Tender Specification	Point raised by Vendor	CBRI team Clarification	Response
3C : X-ray power; 10 Watts or More	Explicitly allow X-ray power of 50W or higher.	The tender specifications are made looking at all major suppliers of EDXRF. We understand the importance of X ray power and assure you that the specifications are generic in nature. We have also mentioned 10Watts or more so you can quote more wattage system.	No change in tender specification.
Measurement atmosphere : Instrument must have capability to test samples in Air, Vacuum or Helium.	Include vacuum mode alongside helium and air.	The tender specifications are made looking at all major suppliers of EDXRF and we have already included this in the tender specification. If you have both Helium and Vacuum option, you are free to offer it and it will be considered during evaluation.	No change in tender specification.
	Add provisions for camera and collimator functionality.	The tender specifications are made looking at all major suppliers of EDXRF. If you have camera and collimator as standard scope of supply with no commercial addition of price, you are free to offer it and it will be considered during evaluation. You can also offer models without such feature as for our application camera is not essential.	No change in tender specification.

Conclusion: Accordingly, Shimadzu India Pvt. Ltd is free to quote their bid. Hence, the specifications provided in the tender are generic, broad, and non restrictive, ensuring that all eligible vendors can participate without any exclusion.